

# Declaration of Conformity

according to Regulation (EU) 2017/745 on medical devices, annex IV dated May 2017

*BEE Medic GmbH*  
*SRN DE-MF-000034511*

*Sattlertorstr. 48a*  
*91301 Forchheim*  
*Germany*

*declares under its sole responsibility that the product*

*Brummi*  
*Article No. A01005*  
*Basic UDI: 426241876BRUMMIEW6H*

The Brummi is a tactile feedback device that serves as an additional feedback channel for biofeedback applications.

*Serial No: 6100 - 8000*

to which this declaration relates is in conformity with the following regulation:

*Evaluation procedure according to Regulation (EU) 2017/745 on medical devices, annex IV, IX, X, XI and related laws*

Classification according to Regulation (EU) 2017/745 on medical devices, annex VIII:

*Active medical device class I according to classification rule 13*

This product *fulfils the CS (Common Specifications) and* has been CE-labelled due to the fulfilment of the general safety and performance requirements according to *Regulation (EU) 2017/745 on medical devices, annex.*



**BEE Medic GmbH**

Dr. Bernhard Wandernoth  
Managing Director

A handwritten signature in blue ink, appearing to be 'B. Wandernoth', written over a horizontal line.

Forchheim, 12.01.2024